

Risk of cancer after primary total hip replacement: The influence of bearings, cementation and the material of the stem

A retrospective cohort study of 8,343 patients with 9 years average follow-up from Valdoltra Orthopaedic Hospital, Slovenia

Vesna Levašič, Ingrid Milošev & Vesna Zadnik

To cite this article: Vesna Levašič, Ingrid Milošev & Vesna Zadnik (2018) Risk of cancer after primary total hip replacement: The influence of bearings, cementation and the material of the stem, Acta Orthopaedica, 89:2, 252-252, DOI: [10.1080/17453674.2018.1440456](https://doi.org/10.1080/17453674.2018.1440456)

To link to this article: <https://doi.org/10.1080/17453674.2018.1440456>



© 2018 The Author(s). Published by Taylor & Francis on behalf of the Nordic Orthopedic Federation.



Published online: 05 Apr 2018.



Submit your article to this journal [↗](#)



Article views: 713



View related articles [↗](#)



View Crossmark data [↗](#)



Citing articles: 1 View citing articles [↗](#)

Erratum

Risk of cancer after primary total hip replacement: The influence of bearings, cementation and the material of the stem

A retrospective cohort study of 8,343 patients with 9 years average follow-up from Valdoltra Orthopaedic Hospital, Slovenia

Vesna LEVAŠIČ^{1,2}, Ingrid MILOŠEV^{1,3}, and Vesna ZADNIK⁴

¹ Valdoltra Orthopaedic Hospital, Ankaran; ² University of Ljubljana, Faculty of Medicine, Ljubljana, ³ Jožef Stefan Institute, Ljubljana; ⁴ Institute of Oncology Ljubljana, Ljubljana, Slovenia

Correspondence: VZadnik@onko-i.si

ACTA ORTHOP 2018; DOI 10.1080/17453674.2018.1431854

Table 1 had some missing zeros and therefore some subgroups (Non-cemented and Ti-alloy) did not add up. A corrected version is presented below:

Table 1. Number of patients in sub-cohorts (bearing surface between head and cup, cementation, material of the stem) by age at operation

Group Subgroup	Age group						Total
	0–40	40–49	50–59	60–69	70–79	≥ 80	
All	131	591	1,631	3,059	2,664	267	8,343
Bearing							
MoM	12	43	141	124	18	0	338
MoP	39	233	763	2,128	2,481	265	5,909
CoC	69	274	535	407	38	0	1,323
CoP	11	41	192	400	127	2	773
Use of bone cement							
Non-cemented	119	580	1,596	2,910	1,713	48	6,966
Cemented	12	11	35	149	951	219	1,377
Stem material							
Ti-alloy	122	585	1,610	2,980	1,940	79	7,316
CoCr-alloy	9	5	21	67	586	162	850
Fe-alloy	0	1	0	12	138	26	177